Inter-laboratory Quality Control Material Data Sheet

QC Material ID: 14143-QC-PR12866

Product Type: Treated Timber (CCA)

Issue Date: 22 May 2026

Expiry Date: 30 September 2027

Test Name	n Values	Assigned Value	Assigned Range
Arsenic (%m/m)	10	0.3769	0.3020 - 0.4518
Chromium (%m/m)	6 *	0.3646	0.3031 - 0.4262
Copper (%m/m)	8	0.1960	0.1538 - 0.2382

** n Values < 8 are indicative only and should be used with caution*

Homogeneity: This QC material has been assessed as homogeneous for use in an Inter-Laboratory Proficiency Study

Stability: An expiry period consistent with the shelf life of the product, experience, and technical advice has been applied

Values: The *Assigned Value* is the consensus mean from an inter-laboratory proficiency round excluding outliers; *n Values* are the number of results used to calculate the assigned value. The *Assigned Range* is 2 standard deviations of *n Values*.

Methods: Multiple methods were used in the preparation of these assigned values.

Global Proficiency Ltd
10 Bisley Road, Enderley, Hamilton 3214, New Zealand
Unit 2/25 Mareno Road, Tullamarine, VIC 3043, Australia
p +64 7 850 4483 | +61 3 9089 1151
e sales@global-proficiency.com
w www.global-proficiency.com

Version 2