GLOBAL PROFICIENCY
INDEPENDENT TEST VERIFICATION

Reference Material - Data Summary Sheet

GP Batch: 1041-C-1

Product Type: Stg 2 (FO) COW Wet-Blend

Issue Date: 24 March 2026

Expiry Date: 9 February 2027

Test Name	n Labs	n Values	Mean	Uncertainty [1] of Mean	k	Methods
Adenosine (5-Mono) (mg/100g)	8	22	8.29	+/- 0.15	2.08	AOAC 2011.20, AOAC 2011.20 (Mod), GB5413.40 (Mod), In House (HPLC), ISO 20638
Ash (g/100g)	9	23	3.06	+/- 0.05	2.07	AOAC 920.153 OMA, BS 1743 (Mod), GB 5009.4 (Method I), In House (Gravimetric) , NZTM3.4.2 / BS 1743 (Mod), NZTM3.4.4
Cytidine (5-Mono) (mg/100g)	8	21	15.1	+/- 0.2	2.09	AOAC 2011.20, AOAC 2011.20 (Mod), GB5413.40 (Mod), In House (HPLC), ISO 20638
Fat (g/100g)	8	23	24.6	+/- 0.1	2.07	AOAC 932.06 / 989.25 / 945.48B (Mod), GB5009.6 (Method III), In House (Gravime tric Extraction), In House (Mojonnier), ISO 23318/IDF 249, NZTM3.6.3
Free Carnitine (mg/kg)	4	9	275	+/- 10	2.31	AOAC 2015.10, AOAC Vol 91 2008, In House (LC-MS), In House (UV-Vis Spectro)
Guanosine (5-Mono) (mg/100g)	8	20	3.27	+/- 0.07	2.09	AOAC 2011.20, AOAC 2011.20 (Mod), GB5413.40 (Mod), In House (HPLC), ISO 20638
Inosine (5-Mono) (mg/100g)	8	21	4.85	+/- 0.10	2.09	AOAC 2011.20, AOAC 2011.20 (Mod), GB5413.40 (Mod), In House (HPLC), ISO 20638
Moisture 102oC (%)	10	39	2.34	+/- 0.04	2.02	AOAC 925.09 / 926.08 (Mod), AS2300.1. 1, GB5009.3 (Method I), GB5009.3 (Meth od I) Mod, GEA NIRO Moisture A1b (Mod), IDF 26A, In House (Gravimetric), In House (Oven Dried), NZTM3.12.8, NZTM3.12.8 / IDF 26A
myo-Inositol (mg/kg)	7	12	606	+/- 25	2.20	AOAC 2011.18, GB5009.270 (Method II), In House (GC-FID), In House (IC-UV), NZTM3.17.53
Protein (g/100g)	10	33	13.3	+/- 0.1	2.04	AOAC 981.10 OMA (Mod), AOCS Ac4-91 (Mod), GB5009.5 (Method I), In House (Kj eldahl), ISO 8968-1/IDF 20-1
Total Carnitine (mg/kg)	6	14	289	+/- 6	2.16	AOAC 2015.10, AOAC 2015.10 (Mod), GB29989, In House (LC-MS), In House (U V-Vis Spectro)
Uridine (5-Mono) (mg/100g)	8	20	8.25	+/- 0.13	2.09	AOAC 2011.20, AOAC 2011.20 (Mod), GB5413.40 (Mod), In House (HPLC), ISO 20638

Global Proficiency Limited
10 Bisley Road, Enderley, Hamilton 3214, New Zealand
p +64 7 850 4483
e sales@global-proficiency.com
w www.global-proficiency.com

ACCREDITED ILAC REFERENCE MATERIAL PRODUCER

[1] *Uncertainty of Mean is an expanded standard error of the mean (SEM) with a level of confidence of 95%. The coverage factor (k) is the critical value of Student's t distribution with n - 1 degrees of freedom. This uncertainty may only represent the measurement reproducibility of the data used. It does not include all known associated uncertainties and does not suggest that all methods used are statistically equivalent.*

Intended Use: For the verification of analytical methods used to determine attributes of Stg 2 (FO) COW Wet-Blend

Description: 150 g+ packed in a white recyclable container

Homogeneity: This reference material has been assessed as homogeneous.

Stability: An expiry period consistent with the shelf life of the unopened product and/or estimated from stability studies has been applied.

Storage: Store samples frozen. Once opened, samples will be unstable due to moisture absorption.

Instructions for Use: Sample test portion size must be as stated in the method used in your laboratory; modification could invalidate the Reference values. The reference values provided for this product are only valid for a single use.

Reference Values: The Mean is the simple mean excluding outliers; *n labs* is the number of laboratories who contributed results, and *n values* is the number of results used to calculate the mean. The uncertainty of the mean is the range within which the true mean of the population can be expected to lie with 95% confidence. This value is derived in accordance with the procedures described in ISO 33405:2024 "Reference materials - Approaches for characterization and assessment of homogeneity and stability".

Quality Statement: The values shown have been generated from data submitted in an inter-laboratory comparison programme (ILCP). The ILCP and reference material production are accredited to ISO/IEC 17043:2023 and ISO 17034:2016 respectively. Generation of this reference material also includes the requirements of ISO 33401 and 33405, which relate to production, characterisation and final product presentation.

Authorised by: Thomas Kuhn (Global P) - Key Technical Person



Global Proficiency Limited
10 Bisley Road, Enderley, Hamilton 3214, New Zealand
p +64 7 850 4483
e sales@global-proficiency.com
w www.global-proficiency.com